

Cardy Nitrate NO₃⁻ Meter

PRODUCT MANUAL

Item 2305G



Spectrum[®]
Technologies, Inc.

CONTENTS

Introduction	3
Meter Components and Functions	4
Handling Precautions	6
Calibrating Solutions	8
Calibration	9
Measurement Procedure	12
Special Setting Modes	14
Replacing the Sensor	18
Installing and Replacing Batteries	19
Preparing Soil Samples	20
Collecting Tissue Sap	23
Preparing Soil-less Media	24
Specifications	26
Soil Test Kit Accessories	27
Sufficiency Levels	28
Warranty	34

This manual will familiarize you with the features and operation of your Cardy Nitrate meter. Please read this manual thoroughly before using your meter. For customer support or to place an order, contact Spectrum Technologies, Inc between 7:30 a.m. and 5:30 p.m. CST.

Phone: (800) 248-8873 or (815) 436-4440

Fax: (815) 436-4460

E-mail: info@specmeters.com.

Website: www.specmeters.com

INTRODUCTION


Congratulations on the purchase of your Cardy Nitrate (NO_3^-) meter. This manual describes how to use your Cardy meter and how to keep it working accurately for many years. Read it thoroughly in order to make effective use of your meter.

Nitrogen is one of the most important elements required for plant growth. It is a key component in chlorophyll, the pigment in plants that harnesses the sun's energy to convert CO_2 into carbohydrates and makes plants green. Tools that provide increased management of N offer significant economic benefits to plant growers of all types. In addition, applying just the right amount of nitrogen at the right time can help safeguard the environment by reducing the amount of residual soil nitrate (NO_3^-) available to leach into ground water

This self-contained, waterproof, digital meter delivers high quality answers to nitrate level questions in soils, plants, and water-based solutions. The replaceable pen-type sensor makes the measurement of small samples much more convenient. When replacement of the sensor is required, the sensor cartridge snaps in and out of the meter at a touch.

The readout of the measurement value is an LCD display. It has a total display range of 14 - 1400 ppm NO_3^- -N. The display can be set to readout in units of nitrate (NO_3^-) or nitrate-nitrogen (NO_3^- -N).

METER COMPONENTS AND FUNCTIONS




No.	Name	Description
1	Flat sensor	Consists of a liquid junction (electrode A) and response membrane (electrode B). Both electrodes must be covered by sample to enable accurate measurement.
2	Light shield cover	Cover used to shield sensor from light. Because the sensor is affected by light, attach the light shield cover before starting measurement.
3	Lithium batteries	CR2032


No.	Name	Description
4	MEAS button	Used during measurement. Press the button for $\frac{1}{2}$ second to evaluate the stability of the measured value. The screen will lock onto a stable value. If this button is pressed for 5 seconds, the meter will go into the special setting mode (p. 14)
5	ON/OFF button	Press and hold for at least 2 seconds to turn meter ON or OFF.
6	CAL button	Press and hold for at least 2 seconds to calibrate the meter.
7	Waterproofing gasket	Provides waterproof seal for meter.
8	MEAS icon	Flashes during calibration and lights steadily when meter locks onto a measurement.
9	CAL icon	Flashes during measurement and lights steadily when calibration is finished.
10	Battery alarm icon	Lights when batteries are low in power and need to be replaced.
11	Temperature alarm icon	Flashes when the measuring environment temperature is out of range (not between 5° and 40°C).
12	Stability icon	Lights only when the measured value has stabilized.
13	Display	Flashes when the measurement is out of range. When Hi is displayed, the result is too high. Dilute the sample and take measurement again.
14	Measurement unit display	The default setting is “NO ₃ ”. See “Special Setting Modes” (p. 16) for instructions to change modes.

HANDLING PRECAUTIONS

- Neither the meter nor sensor is waterproof by itself. The sensor must be securely mounted on the meter before use.
- Never drop the meter or apply excessive force to it.



- The sensor is made of PVC and may get scratched if not handled with care.




- Do not leave the meter in areas of direct sunlight or high temperature/humidity
- Do not clean the meter with organic solvents.



- When using the meter for the first time, or after several weeks of inactivity, the sensor may be slow to respond. Put some reference solution on the sensor and wait 10 to 60 minutes before switching the meter on.

- Measurement of organic solvents, oils, adhesives, cement, alcohols or concentrated acids (pH less than 2) will shorten the sensor's life.
- Some sample types (such as oily plant sap) may not yield stable values.
- To ensure a waterproof seal, check that the waterproofing gasket is not scratched or dirty and is properly seated in the groove.
- Wash off any calibration solution that comes into contact with hands or exposed skin. If fluid gets in the eyes, rinse them immediately and see a doctor.
- White powder may appear on sensor electrode A (the liquid junction) when using the meter in low-humidity environments. Liquid may appear on the electrode in high humidity environments. The appearance of powder or liquid is normal. Simply wash it off with water before using the meter.
- Sensor B is made of PVC. Because the PVC's plasticizer may seep into the fluid being measured, never drink any fluid that has come in contact with the sensor.
- The batteries provided with the meter are for monitoring and may have a short life.
- When replacing batteries, replace both at once.
- Never dispose of batteries in an open flame or attempt to recharge them. Keep used batteries out of reach of children.
- Replace the batteries when the low voltage alarm icon illuminates. When the voltage is low, you may not be able to turn the meter on.
- The sensors cannot be repaired. They must be replaced when they get damaged or when performance deteriorates due to age.



CALIBRATING SOLUTIONS

NO_3^- (nitrate) versus NO_3^- -N (nitrate-nitrogen)

The sensor of the Cardy meter measures NO_3^- nitrate ion activity similar to the way a pH sensor measures H^+ ion activity. Decide whether you desire the meter LCD display to express the sample concentration as NO_3^- nitrate or NO_3^- -N nitrate-nitrogen. (Note: this is similar to referencing length as 6 in. or ½ ft.). Laboratory analysis and university guidelines are generally expressed as NO_3^- -N nitrate-nitrogen.

To convert NO_3^- -N nitrate-nitrogen to NO_3^- nitrate:

$$\begin{array}{ccc} \text{multiply} & \text{or} & \text{divide} \\ \text{NO}_3^- \text{-N} \times 4.42 = \text{NO}_3^- & & \text{NO}_3^- \text{-N} / .226 = \text{NO}_3^- \end{array}$$

Therefore:

$$2000 \text{ ppm } \text{NO}_3^- \text{ nitrate} = 450 \text{ ppm } \text{NO}_3^- \text{-N nitrate-nitrogen}$$

$$150 \text{ ppm } \text{NO}_3^- \text{ nitrate} = 34 \text{ ppm } \text{NO}_3^- \text{-N nitrate-nitrogen}$$

The following table shows the concentrations of the various calibration standards in each unit of measurement.

Description	ppm NO_3^-	ppm NO_3^- -N
#2311 450 ppm NO_3^- -N *	2000	450
#2312 34 ppm NO_3^- -N *	150	34
#2336 450 ppm NO_3^- -N **	2000	450
#2334L 34 ppm NO_3^- -N **	150	34

* For use with plant sap, water and nutrient solutions.


** For use with soil samples (contains aluminum sulfate).

CALIBRATION


Selecting the Calibration Mode.

The meter is shipped with the meter in 2-point calibration mode. The calibration mode can be changed as follows (see also “Special Settings Modes” , p. 14)


1. With the meter on, press and hold the **MEAS** button for at least 5 seconds to enter the special setting mode. When all the LCD elements illuminate, release the **MEAS** button so the number 1 appears. If the button is held too long, the LCD will display the word “Unit”. In this case, press and hold the **MEAS** button to get back into measurement mode and repeat this step.



2. Press and hold the **CAL** button until the **CAL** icon and the number 3 appear. If you pass number 3, press and hold the **CAL** button until it cycles around to number 3




3. Press the **MEAS** button to displays the current calibration mode (1- or 2-point). Press the **CAL** button until the desired number of calibration points appears (It will toggle between 1 and 2).




4. Press the **MEAS** button to save the calibration mode. The meter will return to standard measurement mode.

One-Point Calibration (ppm NO₃⁻ - N mode)


1. Press and hold the ON/OFF button for, at least, two seconds to power the unit on.



2. Open the light shield cover and add enough of the 450 ppm NO₃⁻ - N reference solution to cover the space between electrodes A and B. Rinsing the sensor pad with reference solution beforehand will increase the accuracy of the calibration.



3. Close the light shield cover and press the CAL button until the CAL icon flashes (a minimum of 2 seconds).




4. When the calibration has been finalized, the CAL icon will stop flashing. The reference solution concentration will be displayed along with the 😊 icon. The meter has not properly calibrated if the CAL icon remains flashing or if the **ERR** symbol is displayed on the LCD. In this case, confirm you have used the correct calibration solution, wash the sensor thoroughly and try to re-calibrate the sensor.

5. Wash the sensor with tap water and remove any adhering droplets before taking measurements.


6. Press the **MEAS** button to return to measurement mode.

Two-Point Calibration (ppm NO₃⁻ - N mode)

1. Press and hold the ON/OFF button for, at least, two seconds to power the unit on.



2. Add enough of the 34 ppm NO₃⁻-N reference solution to cover the space between electrodes A and B. Close the light shield cover and press the **CAL** button until the **CAL** icon flashes (a minimum of 2 seconds).



3. When the first calibration has been finalized, the **CAL** icon will stop flashing. The first calibration must be finalized before proceeding to the second calibration.

4. Repeat step 2 with the 450 ppm NO₃⁻-N reference solution.

5. When the second calibration has been finalized, the **CAL** icon will stop flashing and the reference solution concentration will be displayed. The meter has not properly calibrated if the **CAL** icon remains flashing or if "ERR" is displayed on the LCD. In this case, confirm you have used the correct calibration solution, wash the sensor thoroughly and try to re-calibrate the sensor.

If the two calibrations fail but the displayed reference solution concentration is correct, the sensor should be replaced.

6. Wash the sensor with tap water and remove any adhering droplets before taking measurements.


7. Press the **MEAS** button to return to measurement mode.

MEASUREMENT PROCEDURE

Measurement on a flat surface

Important: Do not make measurements with the meter (sensor) in direct sunlight.

1. Open the light shield cover and add enough of the sample to cover the space between electrodes A and B.




2. Close the light shield cover



3. When the stability icon ☺ appears, the LCD will display the current nitrate concentration. To lock in a reading, press the **MEAS** button. The **MEAS** icon will flash to indicate that measurement has begun.



4. When the measurement is complete, the **MEAS** icon will cease flashing and the displayed value will be locked.




5. Press the **MEAS** button again to unlock the LCD and return to dynamic reading mode. After rinsing the sensor, steps 1 - 4 can then be repeated.

*Measurements can be made without pressing the **MEAS** button. In this case, the meter will remain in dynamic measurement mode.*


Submerged Measurement

1. Open the light shield cover's sliding cap. Submerge the sensor in the sample and shake it gently 2 or 3 times
2. When the stability icon ☺ appears, proceed as with flat surface measurements (p. 12).



Scoop Measurement

1. Open the light shield cover's sliding cap. Submerge the sensor in the sample and shake it gently 2 or 3 times. Scoop up some of the sample onto the meter. Place the meter on a flat surface.
2. When the stability icon ☺ appears, proceed as with flat surface measurements (p. 12).



Temperature Measurement

The meter has an internal temperature sensor. A reading can be taken in special settings mode 2 (see p. 15).


SETTING SPECIAL MODES

The Cardy nitrate meter has 6 special modes that can be modified at the user's discretion. The modes are:

1. Set measurement units
2. Temperature measurement
3. Set number of calibrations
4. Set slope coefficient
5. Set offset coefficient
6. Voltage measurement


Entering Special Setting Mode

Press and hold the **MEAS** button for, at least, 5 seconds while the meter is in measurement mode. Release the **MEAS** button as soon as you see all the LCD elements illuminated.



Full display screen

The LCD will then transition to special setting mode starting with the screen for mode 1. If the **MEAS** button is not released immediately after seeing the full display, the meter will enter the “Set measurement units” mode and display the word “Unit”. In this case, press and hold the **MEAS** button to return to measurement mode and start over.




Sample mode selection screen

To scroll to the next mode, press the **CAL** button for half a second. If the **CAL** button is held down, the meter will cycle through the modes sequentially.

Once in the desired mode selection screen, press and hold the **MEAS** button for half a second to modify that configuration. Pressing and holding the **MEAS** button will take you out of special setting mode. See the following pages for details on each mode.

Mode 1 (Set measurement units)




From mode selection screen 1, press the **MEAS** button for half a second. From here, you can configure which units will be displayed. Press the **CAL** button for half a second to cycle to the desired option. The choices are:

ppm NO₃⁻ - N, ppm NO₃⁻, mg/L NO₃⁻ - N, and mg/L NO₃⁻

Press and hold the **MEAS** button to return to measurement mode.


Mode 2 (Temperature measurement)



From mode selection screen 2, press the **MEAS** button for half a second. This is not a mode configuration screen. Instead, in this mode, the meter displays the reading of the internal temperature sensor (in Celsius). The measurement precision is not guaranteed and should only be used as an estimate.

Press and hold the **MEAS** button to return to measurement mode.


Mode 3 (Set number of calibrations)



From mode selection screen 3, press the **MEAS** button for half a second. From here, you can configure the number of solutions that will be used to calibrate the meter. The options are 1- and 2-point calibration. Press the **CAL** button for half a second to cycle to the desired option.

Press and hold the **MEAS** button to return to measurement mode.


Mode 4 (Set slope coefficient)



From mode selection screen 4, press the **MEAS** button for half a second. From here, you can set a multiplicative constant (0.1 to 9.9) that will be multiplied by the reading. This, along with the offset setting (mode 5), can be used to modify the output value to suit a specific application such as dilute or concentrated solutions. The default setting is 1.00.

Press and hold the **MEAS** button to return to measurement mode.


Mode 5 (Set offset coefficient)



From mode selection screen 5, press the **MEAS** button for half a second. From here, you can set an additive constant (-1000 to 1000) that will be added to the reading. This, along with the slope setting (mode 4), can be used to modify the output value to suit a specific application such as dilute or concentrated solutions. The default setting is 0.

Press and hold the **MEAS** button to return to measurement mode.



Mode 6 (Sensor voltage measurement)



From mode selection screen 6, press the **MEAS** button for half a second. This is not a mode configuration screen. Instead, in this mode, the meter displays the reading of the sensor voltage (in mV). Use this mode to evaluate sensor performance or to create your own calibration curve.

Press and hold the **MEAS** button to return to measurement mode.

REPLACING THE SENSOR



Removal




Lift hole **a** off tab **A** and slide the sensor away from the chassis.

Replacement

Slide the sensor onto the body so that tab **A** fits into hole **a**. Take care not to twist the waterproof gasket

Precautions

- As shown here, make sure that the sensor tab is inserted over the copper strip on the main unit. Be careful not to damage the connector




- When removing the sensor, carefully wipe off any drops of sample remaining on the water proof gasket. Do not let water permeate the main body.
- When replacing the sensor or batteries, be sure the unit is powered off.

INSTALLING AND REPLACING BATTERIES


Installation

Slide both batteries into the battery clips as shown such that the positive sides are facing upward.



Replacement

Use a small flat-head screwdriver or other instrument to pry the batteries from the clips as shown. Always replace both batteries at once.



PREPARING SOIL SAMPLES

Sample Collection

Collect at least 15 to 20 core samples from an area not exceeding ten acres by using a Z pattern. Areas having different soil types or management histories should be sampled separately. Sample the top 12 inches of soil. Some universities recommend sampling the top 24 inches in 12 inch increments. Contact your county extension agent for recommendations. Care should be taken to ensure that soil samples are not biased by the presence of rows or bands of fertilizer.

Soil Preparation

Samples should be dried within 24 hours of collection to minimize changes in NO_3^- -N concentrations. Before drying, crumble the soil to avoid large clods that will be difficult to crush when dry. The samples should then be dried by spreading on a thin layer of newspaper at least 3 pages thick and placing overnight in a warm spot with good air movement. Soil will dry in a few hours if placed in a sunny location exposed to the wind. If dried indoors, 24 to 48 hours may be required. Indoor drying time can be reduced with the use of a fan. For oven drying, spread a thin layer of soil on a cookie sheet or pie plate. Place it in an oven set to no more than 250°F with the door slightly ajar. Consider the soil dry when it crumbles rather than compacts under pressure. After drying, crush the soil by using a block of wood or other suitable device. Crush until the soil particles are the size of BB's or smaller. Sift with a flour sifter or other 10 mesh screen. Mix soil thoroughly.

Soil testing of mineral soil requires the Soil Test Kit (item #2330). Be sure you are using the soil standard solutions (item nos. 2334L and 2336) when calibrating the meter for soil tests. Do not use the calibration standards for water/plant sap (See **Calibrating Solutions**, p. 8).

Sample Preparation

1. Measure 2 level measuring spoons (30 ml) full of dry soil into the soil sample cup.
2. Add 2 (30 ml) measuring spoons of the soil extractant to the soil.
3. Mix the soil and the solution by stirring with the spoon for at least 2 minutes, making sure the soil sample is thoroughly mixed. Let stand for 5 minutes
4. Fold a circular filter in half 'twice' and open it up to form a cone. Place it in the soil suspension as far as possible. The filtration will take place from the outside of the filter to the inside.
5. As soon as sufficient filtrate accumulates in the filter, use the small pipette to transfer the soil extract onto the sensor of the Cardy meter.
6. Read the value from the digital display after it has stabilized (30 - 45 sec.). Subtract 34 from the display value. This accounts for the 34 ppm NO_3^- in the extractant solution. For lbs/acre, multiply by 4 for a sampling depth of 12 inches. Note: For soils very low in nitrate, it is possible to get sensor readings less than 34.
7. Rinse sensor and blot dry. Display should read close to "0" with distilled water on it, if not rinse again.

Correction for moist soil

In order to make accurate soil nitrate measurements, the soil must be dry. Nitrate is extracted by mixing a precise amount of dry soil with a measured amount of extracting solution. If the soil is moist, the dilution varies and the result is an understated nitrate test concentration.

When soil nitrate testing is used for **diagnostic** purposes, it may be useful to perform in-field tests using **moist** soil. When performing the rapid test, the content of water in the soil should be taken into account.

The wetter the soil, the greater the tendency for a low result. Precise work requires that the measured nitrate content be multiplied by a correction factor. Table 1 gives the correction factor to be added (in percent) for each type of soil as a function of soil wetness.

Moisture content	Sand	Clay sand	Sandy clay	Clay, Loess
Very dry	+8%	+15%	+17%	+23%
dry	+12%	+20%	+24%	+33%
Med-moist	+12%	+25%	+31%	+43%
moist	+14%	+30%	+38%	+53%
Very moist	+16%	+35%	+45%	+63%

Table 1: Estimated soil nitrate percentage correction for soils at various field moisture levels. Increase the value given by Cardy meter by the percentage shown in the table.

COLLECTING TISSUE SAP

When conducting a test on plant materials, the biggest source of error is due to sampling. This error results when a sample is not representative of the source. Follow these steps to gather and care for your sample:

1.) Do not sample plants which show obvious signs of nutrient deficiency or damage from disease, insects, or chemicals unless these plants are the subject of a study. Plants which have been under stress for a period of time may not give a true picture of the nutrient status of the field.

2.) The leaves or parts of leaves selected should be of the same age and relative position on the plant. The most recently matured leaves should be used. These are the leaves that have stopped expanding in size. The petiole or leaf stem of the leaf or appropriated plant material should be used for the test.

3.) A minimum of 25 petioles or leaves should be collected. This is enough to represent a five to ten acre field if the field is judged to be uniform. Chop up the petioles and mix and sub-sample these pieces for testing. Crops with small, dry petioles, such as strawberries require much larger samples to get enough sap compared to fleshy crops such as tomatoes. Store whole petioles, not leaves, at room temperature for up to 1½ hours or on ice for up to eight hours. Cold petioles should be warmed to room temperature before taking a measurement.

4.) Depending on how succulent the petiole is, use a handheld or hydraulic plant sap press (p. 27) to squeeze sap from the petioles.

PREPARING SOIL-LESS MEDIA

Sample collection

1. Collect sample just before plants are irrigated.
2. Avoid the top layer of media with no roots.
3. Collect root media from the bottom $\frac{2}{3}$ of the pot.
4. Take samples from 10 or more plants distributed in the sample population.
5. When a sufficient amount of root media is collected, mix the sample.

Sample Preparation and Analysis

1:2 Extraction method

Measure a known volume of root media in a beaker or cup (usually 50 to 100 ml or $\frac{1}{4}$ to $\frac{1}{2}$ cup). Fill firmly so it is compressed as it was in the pot. Be consistent when measuring. DO NOT lightly fill or heavily pack the beaker. Place the sample into a cup or beaker.

Add 2 equal volumes of distilled water into the cup, mix the sample and wait 10 minutes. Measure the nitrate after sieving out the large particles. The nitrate level can be read directly from the slurry.

Saturated media extract method

Place 300 to 500 ml (1 to 2 cups) of root media sample in a cup or beaker.

Slowly add distilled water, constantly stirring the sample with a spatula or knife. Add enough distilled water so that the sample behaves like a paste with the surface glistening with water, but with no free water on the surface of the sample.

After 15 minutes, add more water if needed.

Extract the solution from the media using a pipette, Buchner funnel, side arm, flask and vacuum pump, filter bag or sieve. Make any additional measurements (such as EC) using the extracted solution. Table 2 gives a general idea of the nitrate levels to look for in the extracted solution.

Media Type	ppm NO₃⁻-N in extract
Seedlings	40 - 70
Young pot and foliage plants	50 - 90
Pot and bedding plants-growing on	80 - 160
Roses, mums or snapdragons in ground or raised beds	120 - 200
Lettuce and tomatoes in ground beds	125 - 225
Celery transplants	75 - 125

Table 2: Interpretation of Greenhouse Soils: Desirable NO₃⁻-N concentrations in saturated media extract.

SPECIFICATIONS

Principle	Ion electrode method
Readout	4-digit LCD digital display
Reproducibility	±10% of indication value
Measurement Range	14 to 1400 ppm NO ₃ ⁻ - N
Operating Temperature	5 to 40 °C
Operating Humidity	maximum 85% RH
Calibration	1-pt (450 ppm NO ₃ ⁻ -N) or 2-pt (34 and 450 ppm NO ₃ ⁻ -N)
Weight	52g (approx. 1.8 oz)
Power	2 - CR 2032 batteries
Auto Power OFF	30 seconds
Dimensions	165mm x 29mm x 19mm (6.5" x 1.1" x 0.8")
Sample volume	0.3 to 2.0ml

SOIL TEST KIT ACCESSORIES

Description	#2330 Replenishment Kit
Std. Solution 450ppm NO ₃ ⁻ -N (30ml)*	1
Extractant 34ppm NO ₃ ⁻ -N (1 Liter)*	2
Cups - 8oz.	3
Measuring Spoon (29.5 cc)	1
Pipet	1
Filter Papers	30

* Calibration solutions in soil test kits include AlSO₄



Hydraulic Plant Sap Press
(item #2720)



Handheld Plant Sap Press
(item #2725)

PETIOLE NO₃⁻-N SUFFICIENCY LEVELS FOR DRIP-IRRIGATED VEGETABLES

(SOURCE: UC-DAVIS)

Crop	Growth Stage	Petiole NO ₃ ⁻ -N concentration	
		Dry Tissue	Fresh Sap
Broccoli	Mid growth	10,000 - 20,000	1000 - 1600
	Button formation	8000 - 15,000	800 - 1200
	Preharvest	5000 - 8000	600 - 1000
Cabbage	Cupping	*	1200 - 1500
	Early heading	*	1000 - 1200
	Mid heading	*	700 - 900
Cantaloupe	Early flower	12,000 - 15,000	1000 - 1600
	Fruit bulking	8000 - 10,000	800 - 1000
	First harvest	4000 - 6000	700 - 800
Cauliflower	Mid growth	*	1000 - 1600
	Curd development	*	700 - 1000
	Preharvest	*	500 - 800
Celery	Mid growth	7000 - 10,000	600 - 800
	Preharvest	6000 - 10,000	400 - 600
Lettuce	Early head formation	7000 - 10,000	400 - 600
	Preharvest	6000 - 8000	300 - 500
Onion	Bulbs 0.5-1.5 in.	*	350 - 500
Pepper	Vegetative growth	7000 - 10,000	900 - 1200
	Early flower/fruit	5000 - 8000	700 - 1000
	Fruit bulking	5000 - 8000	700 - 1000
	Preharvest	5000 - 7000	700 - 900
Potato (Russet Burbank)	Early vegetative	17,000 - 22,000	1300 - 1600
	Mid tuber/bulking	11,000 - 15,000	900 - 1200
	Late tuber/maturation	6000 - 8000	550 - 700
Sweet Corn	Entire season	*	600 - 700
Tomato	Vegetative growth	10,000 - 14,000	700 - 900
	Early flower/fruit	9000 - 12,000	600 - 800
	Fruit bulking	6000 - 8000	500 - 700
	Preharvest	4000 - 7000	400 - 600
Watermelon	Early flower	12,000 - 15,000	1000 - 1600
	Fruit bulking	8000 - 15,000	700 - 900
	Fruit harvest	5000 - 8000	500 - 700

PETIOLE NO₃-N SUFFICIENCY LEVELS

(SOURCE: UNIVERSITY OF FLORIDA)

Crop	Growth Stage	NO ₃ -N (ppm) Fresh Sap
Cucumber	First blossom	800 - 1000
	Fruits 3-inches long	600 - 800
	First harvest	400 - 600
Broccoli & Collards	Six-leaf stage	800 - 1000
	Just prior to harvest	500 - 800
	At first harvest	300 - 500
Summer Squash	First blossom	900 - 1000
	First harvest	800 - 900
Muskmelon	First blossom	1000 - 1200
	Fruits 2-inches long	800 - 1000
	First harvest	700 - 800
Tomato (field)	First buds	1000 - 1200
	First open flowers	600 - 800
	Fruit 1-inch diameter	400 - 600
	Fruit 2-inch diameter	400 - 600
	First harvest	300 - 400
	Second harvest	200 - 400
Bell Pepper	First flower buds	1400 - 1600
	First open flowers	1400 - 1600
	Fruits half-growth	1200 - 1400
	First harvest	800 - 1000
	Second harvest	500 - 800
Eggplant	First fruit (2-inches long)	1200 - 1600
	First harvest	1000 - 1200
	Mid harvest	800 - 1000
Potatoes	Plants 8-inch tall	1200 - 1400
	First open flowers	1000 - 1400
	50% of flowers open	1000 - 1200
	100% of flowers open	900 - 1200
	Tops falling over	600 - 900

PETIOLE NO₃⁻-N SUFFICIENCY LEVELS (CONT.)

(SOURCE: UNIVERSITY OF FLORIDA)

Crop	Growth Stage	NO₃⁻-N (ppm) Fresh Sap
Annual Hill	November	800 - 900
Strawberries (October planting)	December	600 - 800
	January	600 - 800
	February	300 - 500
	March	200 - 500
Watermelon	April	200 - 500
	Vines 6-inches long	1200 - 1500
	Fruit 2-inches long	1000 - 1200
	Fruits half mature	800 - 1000
	First harvest	600 - 800

**PETIOLE NO₃-N SUFFICIENCY LEVELS FOR POTATOES
(SOURCE: UNIV. WISCONSIN-MADISON)**

Optimum range of nitrate-nitrogen concentrations (dry weight and sap basis) in potato petiole at various stages of growth

Growth Stage (days after emergence)	Norkotah, Norland, Atlantic, Kennebec	Shepody, R. Burbank, Snowden	Onaway Superior
30	2.5 - 2.8	2.0 - 2.3	2.3 - 2.5
40	2.3 - 2.5	1.7 - 2.2	2.0 - 2.3
50	1.8 - 2.3	1.2 - 1.6	1.5 - 1.9
60	1.3 - 1.9	0.8 - 1.1	0.9 - 1.2
70	0.8 - 1.1	0.5 - 0.8	0.4 - 0.6
Sap Basis (ppm NO₃⁻-N)			
30	1900 - 2100	1600 - 1800	1800 - 1900
40	1800 - 2000	1600 - 1700	1600 - 1800
50	1400 - 1800	1000 - 1300	1200 - 1500
60	1110 - 1500	700 - 900	500 - 1000
70	700 - 900	500 - 700	400 - 600

Values from the Cardy can be converted to dry tissue calibration by using the equation:

$$\% \text{Dry Weight NO}_3^- \text{-N} = 0.00142 (\text{ppm sap NO}_3^- \text{-N}) - 0.21$$

**PETIOLE NO₃-N SUFFICIENCY LEVELS RUSSET BUR-
BANK POTATOES
(SOURCE: UNIV. OF MINNESOTA)**

Growth Stage	Petiole NO ₃ ⁻ -N (ppm)	
	Dry Tissue	Fresh Sap
Early Vegetative/tuberization	17,000 - 22,000	1300 - 1600
Mid tuber growth/bulking	11,000 - 15,000	900 - 1600
Late tuber growth/maturation	6,000 - 8,000	550 - 700

PETIOLE NO₃-N SUFFICIENCY LEVELS

(SOURCE: MICHIGAN STATE UNIVERSITY)

The following guidelines are based on one year's research results and will be revised as necessary based on future research findings. Readings taken on youngest fully extended petiole.

Carrots

Adequate petiole sap nitrate concentration

Carrot shoulder diameter (in.)	Nitrate-N (ppm)	Nitrate (ppm)
Prior to sizing	750+	3,300+
0.00 - 0.25	550+	2,420+
0.25 - 0.50	450+	1,980+
0.50 - 0.75	300+	1,320+
0.75 - 1.50	250+	1,100+
> 1.50	200+	880+

Celery

Adequate petiole sap nitrate concentration

Weeks after transplant	Nitrate-N (ppm)	Nitrate (ppm)
0 - 5	800+	3,520+
5 - 6	725+	3,190+
6 - 7	650+	2,860+
7 - 8	575+	2,530+
8 - 9	500+	2,200+
9 - 10	425+	1,870+
10 - 11	350+	1,540+
11+	275+	1,210+

Onions

Adequate petiole sap nitrate concentration

Growth Stage	Nitrate-N (ppm)	Nitrate (ppm)
Up to 5 leaves	800+	2,520+
5 to leaves	600+	2,640+
Bulb initiation	300+	1,320+
Bulb bulking	250+	1,110+

PRE-SIDEDRESS NITRATE (PSNT) SOIL TEST INTERPRETATION

University of Tennessee

<17	ppm NO ₃ ⁻ -N Low
17 - 25	ppm NO ₃ ⁻ -N Low
>25	ppm NO ₃ ⁻ -N Low

Rutgers Cooperative Extension

PSNT Soil Test Level (ppm NO ₃ ⁻ -N)	Sidedress N Recommendation
1 - 15	160
16 - 20	120
21 - 25	80
26 - 30	40
31+	0

University of Wisconsin

PSNT result (ppm N)	Soil Potential*	
	Very High/High N/application Rate (lbs/Acre)	Medium/Low
<10	160	120
11 - 12	150	80
13 - 14	125	80
15 - 17	100	40
18 - 20	60	40
>21	0	0

* consult WMEX pub. A2809

Pennsylvania Nitrogen Soil Test Recommendation (Lbs N/Acre)

(Source: Penn State University)

Soil Test Level (ppm NO ₃ ⁻ -N)	Corn Yield Goal				
	100	125	150	175	200
0 - 10	100	130	160	190	220
11 - 15	75	100	125	150	150
16 - 20	50	75	100	125	125
21 - 25	25	50	75	100	100
25+	0	0	0	0	0

Note: Check you county extension office for updates

WARRANTY

This product is warranted to be free from defects in material or workmanship for one year from the date of purchase. During the warranty period Spectrum will, at its option, either repair or replace products that prove to be defective. This warranty does not cover damage due to improper installation or use, lightning, negligence, accident, or unauthorized modifications, or to incidental or consequential damages beyond the Spectrum product. Before returning a failed unit, you must obtain a Returned Materials Authorization (RMA) from Spectrum. Spectrum is not responsible for any package that is returned without a valid RMA number or for the loss of the package by any shipping company.



This equipment has been manufactured for
Spectrum Technologies, Inc.
12360 S. Industrial Dr. East
Plainfield, IL 60585 USA

The Manufacturer's **DECLARATION OF CONFORMITY** is on file at the above address, and certifies conformity to the following:

Model Number: 2305G
Description: Nitrate Meter
Type: Electrical Equipment for Measurement, Control, and Laboratory Use

Directive: 2004/108/EC
Standards: EN 61326-1 (2006), Class B

A handwritten signature in black ink, appearing to read 'Doug Kieffer', is written over the printed name.

Douglas L. Kieffer,
Soil/Water Products Manager

March 4, 2009

Spectrum[®] ***Technologies, Inc.***

**12360 S. Industrial Dr. E
Plainfield IL 60585
(800) 248-8873 or (815) 436-4440
Fax (815) 436-4460
E-Mail: info@specmeters.com
www.specmeters.com**